

| <i>Sub #</i> | <i>Moisture Z-Values</i> | <i>Protein Z-Values</i> | <i>Ash Z-Values</i> | <i>Falling Number Z-Values</i> |
|--------------|------------------------------|-----------------------------|-------------------------|------------------------------------|
| 19 | 1.53 | 0.95 | 0.58 | 0.20 |
| 38 | 1.21 | | | |
| 50 | 1.82 | 0.30 | 0.93 | 2.01 |
| 54 | 0.10 | 0.47 | 0.24 | 1.72 |
| 55 | 1.24 | 0.03 | 0.87 | 0.59 |
| 56 | 0.14 | 0.33 | 0.52 | 0.77 |
| 58 | 0.35 | 0.54 | 0.28 | 0.12 |
| 93 | 0.29 | 0.26 | 0.24 | 0.10 |
| 137 | 0.97 | 0.69 | 0.17 | 0.30 |
| 139 | 0.35 | 0.51 | 0.70 | 1.11 |
| 164 | 0.64 | 0.26 | 1.74 | 0.91 |
| 247 | 0.34 | 0.40 | 2.82 | 0.81 |
| 308 | 0.55 | 1.24 | 1.27 | 1.03 |
| 392 | 0.25 | 0.22 | 1.61 | 0.87 |
| 537 | 0.52 | | | |
| 592 | 0.25 | 0.95 | 0.06 | 0.51 |
| 593 | 0.25 | 0.36 | 1.27 | |
| 623 | 2.10 | 1.24 | 0.74 | 0.28 |
| 657 | 0.75 | 0.07 | 0.70 | 0.34 |
| 660 | 1.32 | 0.08 | 0.17 | 1.68 |
| 687 | 1.12 | 0.51 | 0.23 | 1.01 |
| 765 | 0.05 | 2.41 | 1.03 | 0.63 |
| 769 | 0.07 | 0.15 | 0.18 | 0.87 |
| 771 | 0.05 | 0.14 | 0.40 | |
| 775 | 1.26 | 0.37 | 1.39 | 0.36 |
| 902 | 2.24 | 0.00 | 0.28 | 1.38 |
| 1052 | 1.14 | 2.77 | 0.12 | 1.50 |
| 1326 | 0.55 | 0.07 | 0.52 | 0.36 |
| 1332 | 1.42 | 0.11 | 0.01 | 0.32 |
| 1352 | 0.55 | 1.21 | 0.69 | 0.65 |
| 1354 | 1.14 | * 14.81 | 0.05 | 0.24 |
| 1427 | 0.25 | 0.26 | 0.52 | 0.87 |
| 1456 | 1.05 | 0.22 | 0.18 | 1.50 |
| 1460 | 0.81 | 1.89 | 0.87 | 0.99 |
| 1472 | 0.82 | 0.22 | 2.19 | 1.62 |
| 1493 | 1.24 | 0.22 | 1.03 | 1.15 |
| 1529 | 0.94 | 0.87 | 0.70 | 1.32 |
| 1531 | 0.43 | 2.45 | 0.70 | |
| 1534 | 1.35 | 2.23 | * 5.02 | |
| 1585 | 0.25 | 0.03 | 1.27 | 1.42 |

| <i>Sub #</i> | <i>Moisture Z-Values</i> | <i>Protein Z-Values</i> | <i>Ash Z-Values</i> | <i>Falling Number Z-Values</i> |
|--------------|------------------------------|-----------------------------|-------------------------|------------------------------------|
| 1642 | 0.64 | 1.31 | 0.86 | 0.75 |
| 1662 | 2.25 | 0.24 | 0.06 | 0.12 |
| 1691 | 0.26 | 0.40 | * 14.30 | |
| 1694 | 0.64 | | 1.61 | 0.79 |
| 1711 | 0.29 | 0.59 | 1.49 | |
| N | 45 | 41 | 41 | 37 |
| Mean | 0.79 | 0.67 | 0.76 | 0.84 |
| Min | 0.05 | 0.00 | 0.01 | 0.10 |
| Max | 2.25 | 2.77 | 2.82 | 2.01 |

* Not included in analytical mean

Z-Values

Each individual z-value represents the decimal number of standard deviations by which an analytical result differs from the "true value", as represented by the mean. The minimum or "perfect" z-value is thus 0.00.

Proficiency in any one analysis over a year's time (11 or 12 monthly, 6 bimonthly, or 4 quarterly results) is determined by their mean z-value plus a penalty for each outlier (*) reported, if any. Proficiency in a series is determined by the mean of the mean z-values (including penalties, if any) for the specified principal analyses in that series.

In general, z-values of less than 2.00, consistently maintained and thus averaging less than 2000 over a year for a series (including outlier penalties, if any), are considered to represent satisfactory accuracy and precision. On the same basis, values of less than 1.00 consistently maintained represent outstanding accuracy and precision.

A detailed description of this rating system is available upon request from AACC headquarters.

CONFIDENTIALITY NOTICE: The data contained herein are confidential and intended for subscriber use only. Data from other subscribers are included for comparison purposes only.